

Estradiol ELISA kit I Human Saliva

Ref: SA-E-6200

This ELISA kit is designed to measure free estradiol levels in human saliva, supporting both research and diagnostic efforts in areas such as menstrual cycle monitoring, pregnancy assessment, menopause evaluation, endocrine disorder investigation, and the study of hormone-dependent cancers.

Sample type	Saliva
Capacity	96 tests
Sensitivity	0.716 pg/mL
Range	0/1 - 100 pg/mL
Assay time	3h
Reactivity	Human

INFORMATIONS

Product name	Estradiol Saliva ELISA Free
Description	<i>In vitro</i> diagnostics / Research enzyme immunoassay (ELISA) for the quantitative determination of active free estradiol in human saliva.
Labels	IVD (EU only) or RUO, CE
Format	96-well plate
Samples	Saliva
Sample volume	100 µL
Reactivity	Human
Standard range	0/1 - 100 pg/mL
Sensitivity	0.716 pg/mL
Specificity	No significant cross-reactivity was observed with Estradiol analogs.
Assay time	3h

PROTOCOLS

Product storage	Store at 2-8°C for up to 6 months
Sample collection & storage	Saliva: collect 5 samples within a period of 2 hours (multiple sampling), before meal. Store samples at 2-8°C for up to one week. For longer periods, store at -20°C.
Sample preparation	Sample incubation (30 min), Enzymatic conjugation (120 min)
ELISA	Incubation, revelation and read steps (30 min)
Detailed protocol	IVD: Download instructions for use RUO: Download instructions for use

Product pictures



Estradiol ELISA Saliva SA-E-6200

Estradiol ELISA Saliva SA-E-6200

Contact information

Immusmol
229 Cours de l'Argonne
33 000 Bordeaux - France
Tel: +33 (0) 5 6431 1170
www.immusmol.com

To order, review, ask for technical support, visit product page at:

<https://www.immusmol.com/shop/estradiol-elisa-kit-human-saliva/>